

## CONSOLIDATED TEST RESULTS SUMMARY

Please see the following pages for full test results.

<b>BULK SKU</b>	<b>BATCH #</b>	<b>LOQ: Limit Of Quantitation</b>	
<b>PRODUCT NAME</b>	<b>SERVING SIZE</b>	<b>LOD: Limit Of Detection</b>	
<b>LABORATORY :</b>	<b>OREGON ACCREDITATION: OR100028</b>	1 g = 10 <sup>-3</sup> kg = 10 <sup>3</sup> mg = 10 <sup>6</sup> µg 1 mg/kg = 1 ppm = 1000 ppb	
POTENCY	PER SERVING	PER GRAM	Percent
Cannabidiol (CBD)	mg/serving	mg/g	%
Total THC (d9-THC, THCA)	mg/serving	mg/g	%
Cannabigerol (CBG)	mg/serving	mg/g	%
Cannabinol (CBN)	mg/serving	mg/g	%
Cannabichromene (CBC)	mg/serving	mg/g	%
Tetrahydrocannabinolic Acid (THCA)	mg/serving	mg/g	%
Delta-9-THC (d9-THC)	mg/serving	mg/g	%
Delta-8-THC (d8-THC)	mg/serving	mg/g	%
HEAVY METALS	PER SERVING	PER GRAM	REGULATORY ACTION LEVEL
Arsenic	µg/serving	µg/g	10 µg/day <sup>[1]</sup>
Cadmium	µg/serving	µg/g	4.1 µg/day <sup>[1]</sup>
Lead	µg/serving	µg/g	3.5 µg/day <sup>[2]</sup>
Mercury	µg/serving	µg/g	2 µg/day <sup>[1]</sup>
PESTICIDES	REGULATORY ACTION LEVEL		
None of the other 59 pesticides tested found above limit of detection in the sample.			10 ppb <sup>[1]</sup>
RESIDUAL SOLVENTS	Results	REGULATORY ACTION LEVEL	
Ethanol		50,000 mg/day	
Heptane		50,000 mg/day	
None of the 34 residual solvents tested found above limit of quantitation in the sample.			



1. American Herbal Pharmacopoeia. (2014). Cannabis Inflorescence: Standards of Identity, Analysis, and Quality Control. Washington DC: AHP.  
2. US Food and Drug Administration. (2019). Lead in Food, Foodwares, and Dietary Supplements. Washington DC: FDA. US Food and Drug Administration. (2019). Lead in Food, Foodwares, and Dietary Supplements. Washington DC: FDA.



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794



**Report Number:** 22-009252/D003.R000  
**Report Date:** 08/10/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/04/22 10:49

**Customer:** Etz Hayim Holdings  
**Product identity:** PROC-EG30-CBD  
**Client/Metric ID:** .  
**Laboratory ID:** 22-009252-0001

### Summary

**Potency:**

Analyte	Result (%)		
CBD	> 98.0		<ul style="list-style-type: none"> <li><span style="color: red;">●</span> CBD</li> <li><span style="color: blue;">●</span> CBDV</li> </ul>
CBDV	0.609	THC-Total <LOQ (Reported in percent of total sample)	

**Residual Solvents:**

*All analytes passing and less than LOQ.*



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794



**Report Number:** 22-009252/D003.R000  
**Report Date:** 08/10/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/04/22 10:49

**Customer:** Etz Hayim Holdings  
 16427 NE Airport Way  
 PORTLAND 97230  
 United States of America (USA)

**Product identity:** PROC-EG30-CBD

**Client/Metric ID:** .

**Sample Date:**

**Laboratory ID:** 22-009252-0001

**Evidence of Cooling:** No

**Temp:** 23 °C

**Relinquished by:** UPS

### Sample Results

Potency	Method: J AOAC 2015 V98-6 (mod) <sup>p</sup>		Units %	Batch: 2206648	Analyze: 8/6/22 12:22:00 AM
Analyte	As Received	Dry weight	LOQ	Notes	
CBC	< LOQ		0.0764		
CBC-A	< LOQ		0.0764		
CBC-Total	< LOQ		0.143		
CBD	> 98.0		0.764		
CBD-A	< LOQ		0.0764		
CBD-Total	> 98.0		0.831		
CBDV	0.609		0.0764		
CBDV-A	< LOQ		0.0764		
CBDV-Total	0.609		0.143		
CBE	< LOQ		0.0764		
CBG	< LOQ		0.0764		
CBG-A	< LOQ		0.0764		
CBG-Total	< LOQ		0.143		
CBL	< LOQ		0.0764		
CBL-A	< LOQ		0.0764		
CBL-Total	< LOQ		0.143		
CBN	< LOQ		0.0764		
CBT	< LOQ		0.0764		
Δ8-THC	< LOQ		0.0764		
Δ8-THCV	< LOQ		0.0764		
Δ9-THC	< LOQ		0.0764		
exo-THC	< LOQ		0.0764		
THC-A	< LOQ		0.0764		
THC-Total	< LOQ		0.143		
THCV	< LOQ		0.0764		
THCV-A	< LOQ		0.0764		
THCV-Total	< LOQ		0.143		
<b>Total Cannabinoids</b>	> 98.0				



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794



**Report Number:** 22-009252/D003.R000  
**Report Date:** 08/10/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/04/22 10:49

Solvents											Method: Residual Solvents by GC/MS <sup>b</sup>					Units µg/g		Batch 2206739		Analyze 08/10/22 09:08 AM				
Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes													
1,4-Dioxane	< LOQ	380	100	pass		2-Butanol	< LOQ	5000	200	pass														
2-Ethoxyethanol	< LOQ	160	30.0	pass		2-Methylbutane (Isopentane)	< LOQ		200															
2-Methylpentane	< LOQ		30.0			2-Propanol (IPA)	< LOQ	5000	200	pass														
2,2-Dimethylbutane	< LOQ		30.0			2,2-Dimethylpropane (neo-pentane)	< LOQ		200															
2,3-Dimethylbutane	< LOQ		30.0			3-Methylpentane	< LOQ		30.0															
Acetone	< LOQ	5000	200	pass		Acetonitrile	< LOQ	410	100	pass														
Benzene	< LOQ	2.00	1.00	pass		Butanes (sum)	< LOQ	5000	400	pass														
Cyclohexane	< LOQ	3880	200	pass		Ethanol	< LOQ		200															
Ethyl acetate	< LOQ	5000	200	pass		Ethyl benzene	< LOQ		200															
Ethyl ether	< LOQ	5000	200	pass		Ethylene glycol	< LOQ	620	200	pass														
Ethylene oxide	< LOQ	50.0	20.0	pass		Hexanes (sum)	< LOQ	290	150	pass														
Isopropyl acetate	< LOQ	5000	200	pass		Isopropylbenzene (Cumene)	< LOQ	70.0	30.0	pass														
m,p-Xylene	< LOQ		200			Methanol	< LOQ	3000	200	pass														
Methylene chloride	< LOQ	600	60.0	pass		Methylpropane (Isobutane)	< LOQ		200															
n-Butane	< LOQ		200			n-Heptane	< LOQ	5000	200	pass														
n-Hexane	< LOQ		30.0			n-Pentane	< LOQ		200															
o-Xylene	< LOQ		200			Pentanes (sum)	< LOQ	5000	600	pass														
Propane	< LOQ	5000	200	pass		Tetrahydrofuran	< LOQ	720	100	pass														
Toluene	< LOQ	890	100	pass		Total Xylenes	< LOQ		400															
Total Xylenes and Ethyl benzene	< LOQ	2170	600	pass																				



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 22-009252/D003.R000  
**Report Date:** 08/10/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/04/22 10:49

These test results are representative of the individual sample selected and submitted by the client.

**Abbreviations**

**Limits:** Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

**Limit(s) of Quantitation (LOQ):** The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

<sup>p</sup> = ISO/IEC 17025:2017 accredited method.

**Units of Measure**

µg/g = Microgram per gram

% = Percentage of sample

% wt = µg/g divided by 10,000

Approved Signatory

Derrick Tanner  
General Manager



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794



**Report Number:** 22-009252/D003.R000  
**Report Date:** 08/10/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/04/22 10:49



12423 NE Whitaker Way Portland OR 97230 p.503-254-1794

**Cannabis Chain of Custody Record**

ORELAP ID: OR100028

Field ID		Date/Time Collected	Pesticides - DR 59 compounds	Pesticide Multi-Residue - 379 compounds	Potency	Residual Solvents	Water Activity	Moisture	Terpenes	Micro: Yeast and Mold	Micro: E. Coli and Total Coliform	Heavy Metals	Mycotoxins	Other	Matrix	Weight	Serving size for edibles	Comments/Metric ID
PROC-EG30-CBD		08.03.22			1	1									ISOLATE			

Purchase Order Number:  
 Project Number:  
 Project Name:  
 Report Instructions:  
 Send to State - METRC  
 Email Final Results:  
 Fax Final Results  
 Cash/Check/CC/Net 30  
 Other:

Collected By:	Ref: [Redacted]	Date:	Time:	Project:	Lot:	Lab Use Only:
<input type="checkbox"/> Standard (5 day) <input type="checkbox"/> Rush (3-4 day) (1.5x Standard) <input checked="" type="checkbox"/> Priority Rush (2 day) (2x Standard)						Client Alias: Order Number: Proper Container: Sample Condition: Temperature: 23.0°C Shipped Via: UPS Evidence of cooling: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No

SUBMISSION OF SAMPLES WITH TESTING REQUIREMENTS TO PIXIS WILL BE UNDERSTOOD TO BE AN AGREEMENT FOR SERVICES IN ACCORDANCE WITH THE CONDITIONS LISTED ON THE BACK OF THIS FORM  
 Revision: 1.02 Control#: CF023  
 Effective 01/31/2019 Revised 01/31/2019  
[www.pixislabs.com](http://www.pixislabs.com)  
 Page 1 of 2

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Testing in accordance with: OAR 333-007-0410 OAR 333-007-0430



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



Report Number: 22-009252/D003.R000  
Report Date: 08/10/2022  
ORELAP#: OR100028  
Purchase Order:  
Received: 08/04/22 10:49



Document ID: 3177 Revision: 3  
Effective: 04/26/2022  
Page 1 of 1

PACKAGE RECEIVING FORM

Delivery Date:  \_\_\_\_\_  Same as Opened By Date  Unsure

How was the package delivered?

UPS  FEDEX  USPS  DHL  OTHER: \_\_\_\_\_

Tracking Number: 1Z 301 E88 01 9249 7622


		CIRCLE ONE	
1) Was package sealed with no evidence of holes/tampering?		<input checked="" type="radio"/> YES	<input type="radio"/> NO
Further custody seal/tampering notes: _____			
2) Was packing material used?		YES	<input checked="" type="radio"/> NO
If YES: <input type="checkbox"/> PEANUTS <input type="checkbox"/> BUBBLE <input type="checkbox"/> WRAP <input type="checkbox"/> FOAM PAPER			
3) Was a Complete Chain of Custody (COC) received?		<input checked="" type="radio"/> YES	<input type="radio"/> NO
Comment (PT?, Email?): _____			
4) Sample temperature upon arrival?		<u>23.6</u>	°C
5) Evidence of cooling?		YES	<input checked="" type="radio"/> NO
If YES, What kind? <input type="checkbox"/> ICE <input type="checkbox"/> FREEZER PACK <input type="checkbox"/> DRY ICE			
Insulation? <input type="checkbox"/> PLASTIC COOLER <input type="checkbox"/> STYROFOAM <input type="checkbox"/> OTHER: _____			
6) Were sample containers sealed in separate plastic bags/secondary containment?		<input checked="" type="radio"/> YES	<input type="radio"/> NO
7) Did sample containers arrive in good condition?		<input checked="" type="radio"/> YES	<input type="radio"/> NO
If NO: <input type="checkbox"/> LEAKED <input type="checkbox"/> BROKEN <input type="checkbox"/> OTHER: _____			
If NO: Suspect contamination of other samples? <input type="checkbox"/> YES <input type="checkbox"/> NO			
8) Sample labels present?		<input checked="" type="radio"/> YES	<input type="radio"/> NO
9) Do sample labels agree with COC?		<input checked="" type="radio"/> YES	<input type="radio"/> NO
If NO, number of sample containers received: _____			

Sample pre-log location:

R39 R44 F44 R99  CANNA SHELF  FOOD SHELF Other: \_\_\_\_\_

Other Notes:

Received By (initials)





12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794

**Report Number:** 22-009252/D003.R000  
**Report Date:** 08/10/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/04/22 10:49



Revision: 1 Document ID: 7148  
 Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

JAOAC2015 V986 Batch ID: 2206648

Analyte	LCS	Result	Spike	Units	% Rec	Limits	Evaluation	Notes
CBDA	1	0.102	0.100	%	102	80.0 - 120	Acceptable	
CBDV	1	0.111	0.100	%	111	80.0 - 120	Acceptable	
CBE	1	0.0989	0.100	%	98.9	80.0 - 120	Acceptable	
CBDA	1	0.0981	0.100	%	98.1	90.0 - 110	Acceptable	
CBSA	1	0.0977	0.100	%	97.7	80.0 - 120	Acceptable	
CBS	1	0.103	0.100	%	103	80.0 - 120	Acceptable	
CB	1	0.102	0.100	%	102	90.0 - 110	Acceptable	
THCV	1	0.0978	0.100	%	97.8	80.0 - 120	Acceptable	
δ8THCV	1	0.0996	0.100	%	99.6	80.0 - 120	Acceptable	
THCVA	1	0.0970	0.100	%	97.0	80.0 - 120	Acceptable	
CBN	1	0.103	0.100	%	103	90.0 - 110	Acceptable	
exo-THC	1	0.0964	0.100	%	96.4	80.0 - 120	Acceptable	
δ9THC	1	0.106	0.100	%	106	90.0 - 110	Acceptable	
δ8THC	1	0.0947	0.100	%	94.7	90.0 - 110	Acceptable	
CB	1	0.0907	0.100	%	90.7	80.0 - 120	Acceptable	
CB	1	0.0974	0.100	%	97.4	80.0 - 120	Acceptable	
THCA	1	0.0951	0.100	%	95.1	90.0 - 110	Acceptable	
CBGA	1	0.101	0.100	%	101	80.0 - 120	Acceptable	
CBGA	1	0.104	0.100	%	104	80.0 - 120	Acceptable	
CB	1	0.0903	0.100	%	90.3	80.0 - 120	Acceptable	

Method Blank

Analyte	Result	LOQ	Units	Limits	Evaluation	Notes
CBDA	<LOQ	0.077	%	< 0.077	Acceptable	
CBDV	<LOQ	0.077	%	< 0.077	Acceptable	
CBE	<LOQ	0.077	%	< 0.077	Acceptable	
CBDA	<LOQ	0.077	%	< 0.077	Acceptable	
CBSA	<LOQ	0.077	%	< 0.077	Acceptable	
CBS	<LOQ	0.077	%	< 0.077	Acceptable	
CB	<LOQ	0.077	%	< 0.077	Acceptable	
THCV	<LOQ	0.077	%	< 0.077	Acceptable	
δ8THCV	<LOQ	0.077	%	< 0.077	Acceptable	
THCVA	<LOQ	0.077	%	< 0.077	Acceptable	
CBN	<LOQ	0.077	%	< 0.077	Acceptable	
exo-THC	<LOQ	0.077	%	< 0.077	Acceptable	
δ9THC	<LOQ	0.077	%	< 0.077	Acceptable	
δ8THC	<LOQ	0.077	%	< 0.077	Acceptable	
CB	<LOQ	0.077	%	< 0.077	Acceptable	
CB	<LOQ	0.077	%	< 0.077	Acceptable	
THCA	<LOQ	0.077	%	< 0.077	Acceptable	
CBGA	<LOQ	0.077	%	< 0.077	Acceptable	
CBGA	<LOQ	0.077	%	< 0.077	Acceptable	
CB	<LOQ	0.077	%	< 0.077	Acceptable	

Abbreviations

ND - None Detected at or above MRL  
 RPD - Relative Percent Difference  
 LOQ - Limit of Quantitation

Units of Measure:

% - Percent





12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794

**Report Number:** 22-009252/D003.R000  
**Report Date:** 08/10/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/04/22 10:49



Revision: 1 Document ID: 7148  
 Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

JAOAC2015 V986		Batch ID: 2206648						
Sample Duplicate		Sample ID: 22-008854000101						
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDVA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBDV	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBE	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBD	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBSA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBS	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBD	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
THCV	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
d8THCV	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
THCVA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBN	>98	>98	0.077	%	NA	< 20	Acceptable	
exo-THC	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
d9THC	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
d8THC	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBL	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CB	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
THCA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBGA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CBLA	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	
CB	<LOQ	<LOQ	0.077	%	NA	< 20	Acceptable	

**Abbreviations**

ND - None Detected at or above MRL  
 RPD - Relative Percent Difference  
 LOQ - Limit of Quantitation

**Units of Measure:**

% - Percent



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794

Report Number: 22-009252/D003.R000  
 Report Date: 08/10/2022  
 ORELAP#: OR100028  
 Purchase Order:  
 Received: 08/04/22 10:49



Revision: Document ID:  
 Legacy ID: Effective:

Laboratory Quality Control Results

Residual Solvents				Batch ID: 2206739					
Method Blank				Laboratory Control Sample					
Analyte	Result	LOQ	Notes	Result	Spike	Units	% Rec	Limits	Notes
Propane	ND	< 200		515	572	µg/g	90.0	60 - 120	
Isobutane	ND	< 200		701	731	µg/g	95.9	60 - 120	
Butane	ND	< 200		692	731	µg/g	94.7	60 - 120	
2,2-Dimethylpropane	ND	< 200		912	936	µg/g	97.4	60 - 120	
Methanol	ND	< 200		1460	1650	µg/g	88.5	60 - 120	
Ethylene Oxide	ND	< 30		51	56.2	µg/g	90.7	60 - 120	
2-Methylbutane	ND	< 200		1330	1620	µg/g	82.1	60 - 120	
Pentane	ND	< 200		1370	1610	µg/g	85.1	60 - 120	
Ethanol	ND	< 200		1450	1620	µg/g	89.5	70 - 130	
Ethyl Ether	ND	< 200		1440	1600	µg/g	90.0	60 - 120	
2,2-Dimethylbutane	ND	< 30		149	167	µg/g	89.2	60 - 120	
Acetone	ND	< 200		1460	1620	µg/g	90.1	60 - 120	
2-Propanol	ND	< 200		1520	1610	µg/g	94.4	60 - 120	
Ethyl Formate	ND	< 500		1350	1620	µg/g	83.3	70 - 130	
Acetonitrile	ND	< 100		580	635	µg/g	91.3	60 - 120	
Methyl Acetate	ND	< 500		1480	1630	µg/g	90.8	70 - 130	
2,3-Dimethylbutane	ND	< 30		150	177	µg/g	84.7	60 - 120	
Dichloromethane	ND	< 60		438	498	µg/g	88.0	60 - 120	
2-Methylpentane	ND	< 30		147	166	µg/g	88.6	60 - 120	
MTBE	ND	< 500		1460	1600	µg/g	91.3	70 - 130	
3-Methylpentane	ND	< 30		158	175	µg/g	90.3	60 - 120	
Hexane	ND	< 30		156	174	µg/g	89.7	60 - 120	
1-Propanol	ND	< 500		1370	1620	µg/g	84.6	70 - 130	
Methylethylketone	ND	< 500		1400	1600	µg/g	87.5	70 - 130	
Ethyl acetate	ND	< 200		1450	1610	µg/g	90.1	60 - 120	
2-Butanol	ND	< 200		1440	1620	µg/g	88.9	60 - 120	
Tetrahydrofuran	ND	< 100		445	507	µg/g	87.8	60 - 120	
Cyclohexane	ND	< 200		1430	1610	µg/g	88.8	60 - 120	
2-methyl-1-propanol	ND	< 500		1330	1640	µg/g	81.1	70 - 130	
Benzene	ND	< 1		4.39	5.22	µg/g	84.1	60 - 120	
Isopropyl Acetate	ND	< 200		1460	1610	µg/g	90.7	60 - 120	
Heptane	ND	< 200		1430	1610	µg/g	88.8	60 - 120	
1-Butanol	ND	< 500		1330	1610	µg/g	82.6	70 - 130	
Propyl Acetate	ND	< 500		1340	1610	µg/g	83.2	70 - 130	
1,4-Dioxane	ND	< 100		455	508	µg/g	89.6	60 - 120	
2-Ethoxyethanol	ND	< 30		147	165	µg/g	89.1	60 - 120	
Methylisobutylketone	ND	< 500		1260	1610	µg/g	78.3	70 - 130	
3-Methyl-1-butanol	ND	< 500		1160	1600	µg/g	72.5	70 - 130	
Ethylene Glycol	ND	< 200		169	492	µg/g	34.3	60 - 120	Q6
Toluene	ND	< 100		418	497	µg/g	84.1	60 - 120	
Isobutyl Acetate	ND	< 500		1100	1610	µg/g	68.3	70 - 130	Q6
1-Pentanol	ND	< 500		881	1600	µg/g	55.1	70 - 130	Q6
Butyl Acetate	ND	< 500		1230	1610	µg/g	76.4	70 - 130	
Ethylbenzene	ND	< 200		778	980	µg/g	79.4	60 - 120	
m,p-Xylene	ND	< 200		772	985	µg/g	78.4	60 - 120	
o-Xylene	ND	< 200		751	965	µg/g	77.8	60 - 120	
Cumene	ND	< 30		131	168	µg/g	78.0	60 - 120	
Anisole	ND	< 500		1120	1600	µg/g	70.0	70 - 130	
DMSO	ND	< 500		1140	1610	µg/g	70.8	70 - 130	
1,2-dimethoxyethane	ND	< 50		136	165	µg/g	82.4	70 - 130	
Triethylamine	ND	< 500		1250	1620	µg/g	77.2	70 - 130	
N,N-dimethylformamide	ND	< 150		382	481	µg/g	79.4	70 - 130	
N,N-dimethylacetamide	ND	< 150		338	480	µg/g	70.4	70 - 130	
Pyridine	ND	< 50		111	171	µg/g	64.9	70 - 130	Q6
1,2-Dichloroethane	ND	< 1		0.786	1	µg/g	78.6	70 - 130	
Chloroform	ND	< 1		0.797	1	µg/g	79.7	70 - 130	
Trichloroethylene	ND	< 1		0.804	1	µg/g	80.4	70 - 130	



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794



**Report Number:** 22-009252/D003.R000  
**Report Date:** 08/10/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/04/22 10:49

Revision: Document ID:  
 Legacy ID: Effective:

QC - Sample Duplicate			Sample ID: 22-008977-0001					
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Accept/Fail	Notes
Propane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Isobutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Butane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylpropane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Methanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylene Oxide	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Methylbutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Pentane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Ether	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Acetone	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Propanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Formate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Acetonitrile	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Methyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
2,3-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Dichloromethane	ND	ND	60	µg/g	0.0	< 20	Acceptable	
2-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
MTBE	ND	ND	500	µg/g	0.0	< 20	Acceptable	
3-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Hexane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
1-Propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Methyl ethyl ketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethyl acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Butanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Tetrahydrofuran	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Cyclohexane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-methyl-1-propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Benzene	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Isopropyl Acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Heptane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
1-Butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Propyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,4-Dioxane	ND	ND	100	µg/g	0.0	< 20	Acceptable	
2-Ethoxyethanol	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Methylisobutylketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
3-Methyl-1-butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylene Glycol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Toluene	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Isobutyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1-Pentanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Butyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylbenzene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
m,p-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
o-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Cumene	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Anisole	ND	ND	500	µg/g	0.0	< 20	Acceptable	
DMSO	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,2-dimethoxyethane	ND	ND	50	µg/g	0.0	< 20	Acceptable	
Triethylamine	ND	ND	500	µg/g	0.0	< 20	Acceptable	
N,N-dimethylformamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
N,N-dimethylacetamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
Pyridine	ND	ND	50	µg/g	0.0	< 20	Acceptable	
1,2-Dichloroethane	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Chloroform	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Trichloroethylene	ND	ND	1	µg/g	0.0	< 20	Acceptable	

**Abbreviations**

ND - None Detected at or above MRL  
 RPD - Relative Percent Difference  
 LOQ - Limit of Quantitation  
 Q6 - Quality control outside QC limits. Data acceptable based on remaining QC.

**Units of Measure:**

µg/g - Microgram per gram or ppm



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 22-009252/D003.R000  
**Report Date:** 08/10/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/04/22 10:49





Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitation level raised due to matrix interference.
B	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 22-010282/D002.R000  
**Report Date:** 08/31/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/29/22 09:42

**Customer:** Etz Hayim Holdings  
**Product identity:** PROC-EG30-CBD  
**Client/Metric ID:** .  
**Laboratory ID:** 22-010282-0001

### Summary

-----  
**Pesticides:**

*All analytes passing and less than LOQ.*

**Metals:**

*Less than LOQ for all analytes.*

-----



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 22-010282/D002.R000  
**Report Date:** 08/31/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/29/22 09:42

**Customer:** Etz Hayim Holdings  
16427 NE Airport Way  
PORTLAND 97230  
United States of America (USA)

**Product identity:** PROC-EG30-CBD

**Client/Metric ID:** .

**Sample Date:**

**Laboratory ID:** 22-010282-0001

**Evidence of Cooling:** No

**Temp:** 21.2 °C

**Relinquished by:** USPS

### Sample Results

Pesticides											
Method: AOAC 2007.01 & EN 15662 (mod) <sup>b</sup>											
Units mg/kg Batch 2207341 Analyze 08/31/22 08:34 AM											
Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
Abamectin <sup>×</sup>	< LOQ	0.50	0.250	pass		Acephate <sup>×</sup>	< LOQ	0.40	0.250	pass	
Acequinocyl <sup>×</sup>	< LOQ	2.0	1.00	pass		Acetamiprid <sup>×</sup>	< LOQ	0.20	0.100	pass	
Aldicarb <sup>×</sup>	< LOQ	0.40	0.200	pass		Azoxystrobin <sup>×</sup>	< LOQ	0.20	0.100	pass	
Bifenazate <sup>×</sup>	< LOQ	0.20	0.100	pass		Bifenthrin <sup>×</sup>	< LOQ	0.20	0.100	pass	
Boscalid <sup>×</sup>	< LOQ	0.40	0.200	pass		Carbaryl <sup>×</sup>	< LOQ	0.20	0.100	pass	
Carbofuran <sup>×</sup>	< LOQ	0.20	0.100	pass		Chlorantraniliprole <sup>×</sup>	< LOQ	0.20	0.100	pass	
Chlorfenapyr <sup>×</sup>	< LOQ	1.0	0.500	pass		Chlorpyrifos <sup>×</sup>	< LOQ	0.20	0.100	pass	
Clofentezine <sup>×</sup>	< LOQ	0.20	0.100	pass		Cyfluthrin <sup>×</sup>	< LOQ	1.0	0.500	pass	
Cypermethrin <sup>×</sup>	< LOQ	1.0	0.500	pass		Daminozide <sup>×</sup>	< LOQ	1.0	0.500	pass	
Diazinon <sup>×</sup>	< LOQ	0.20	0.100	pass		Dichlorvos <sup>×</sup>	< LOQ	1.0	0.500	pass	
Dimethoate <sup>×</sup>	< LOQ	0.20	0.100	pass		Ethoprophos <sup>×</sup>	< LOQ	0.20	0.100	pass	
Etofenprox <sup>×</sup>	< LOQ	0.40	0.200	pass		Etoxazole <sup>×</sup>	< LOQ	0.20	0.100	pass	
Fenoxycarb <sup>×</sup>	< LOQ	0.20	0.100	pass		Fenpyroximate <sup>×</sup>	< LOQ	0.40	0.200	pass	
Fipronil <sup>×</sup>	< LOQ	0.40	0.200	pass		Fonicamid <sup>×</sup>	< LOQ	1.0	0.400	pass	
Fludioxonil <sup>×</sup>	< LOQ	0.40	0.200	pass		Hexythiazox <sup>×</sup>	< LOQ	1.0	0.400	pass	
Imazalil <sup>×</sup>	< LOQ	0.20	0.100	pass		Imidacloprid <sup>×</sup>	< LOQ	0.40	0.200	pass	
Kresoxim-methyl <sup>×</sup>	< LOQ	0.40	0.200	pass		Malathion <sup>×</sup>	< LOQ	0.20	0.100	pass	
Metalaxyl <sup>×</sup>	< LOQ	0.20	0.100	pass		Methiocarb <sup>×</sup>	< LOQ	0.20	0.100	pass	
Methomyl <sup>×</sup>	< LOQ	0.40	0.200	pass		MGK-264 <sup>×</sup>	< LOQ	0.20	0.100	pass	
Myclobutanil <sup>×</sup>	< LOQ	0.20	0.100	pass		Naled <sup>×</sup>	< LOQ	0.50	0.250	pass	
Oxamyl <sup>×</sup>	< LOQ	1.0	0.500	pass		Paclobutrazole <sup>×</sup>	< LOQ	0.40	0.200	pass	
Parathion-Methyl <sup>×</sup>	< LOQ	0.20	0.200	pass		Permethrin <sup>×</sup>	< LOQ	0.20	0.100	pass	
Phosmet <sup>×</sup>	< LOQ	0.20	0.100	pass		Piperonyl butoxide <sup>×</sup>	< LOQ	2.0	1.00	pass	
Prallethrin <sup>×</sup>	< LOQ	0.20	0.200	pass		Propiconazole <sup>×</sup>	< LOQ	0.40	0.200	pass	
Propoxur <sup>×</sup>	< LOQ	0.20	0.100	pass		Pyrethrin I (total) <sup>×</sup>	< LOQ	1.0	0.500	pass	
Pyridaben <sup>×</sup>	< LOQ	0.20	0.100	pass		Spinosad <sup>×</sup>	< LOQ	0.20	0.100	pass	
Spiromesifen <sup>×</sup>	< LOQ	0.20	0.100	pass		Spirotetramat <sup>×</sup>	< LOQ	0.20	0.100	pass	
Spiroxamine <sup>×</sup>	< LOQ	0.40	0.200	pass		Tebuconazole <sup>×</sup>	< LOQ	0.40	0.200	pass	
Thiacloprid <sup>×</sup>	< LOQ	0.20	0.100	pass		Thiamethoxam <sup>×</sup>	< LOQ	0.20	0.100	pass	
Trifloxystrobin <sup>×</sup>	< LOQ	0.20	0.100	pass							



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 22-010282/D002.R000  
**Report Date:** 08/31/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/29/22 09:42

Metals

Analyte	Result	Limits	Units	LOQ	Batch	Analyzed Method	Status	Notes
Arsenic	< LOQ	0.200	mg/kg	0.0875	2207329	08/29/22 AOAC 2013.06 (mod.) <sup>p</sup>	pass	
Cadmium	< LOQ	0.200	mg/kg	0.0875	2207329	08/29/22 AOAC 2013.06 (mod.) <sup>p</sup>	pass	
Lead	< LOQ	0.500	mg/kg	0.0875	2207329	08/29/22 AOAC 2013.06 (mod.) <sup>p</sup>	pass	
Mercury	< LOQ	0.100	mg/kg	0.0437	2207329	08/29/22 AOAC 2013.06 (mod.) <sup>p</sup>	pass	





12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 22-010282/D002.R000  
**Report Date:** 08/31/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/29/22 09:42

These test results are representative of the individual sample selected and submitted by the client.

**Abbreviations**

**Limits:** Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

**Limit(s) of Quantitation (LOQ):** The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

Ⓟ = ISO/IEC 17025:2017 accredited method.

\* = TNI accredited analyte.

**Units of Measure**

mg/kg = Milligram per kilogram = parts per million (ppm)

% wt = µg/g divided by 10,000

Approved Signatory

Derrick Tanner  
General Manager



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 22-010282/D002.R000  
**Report Date:** 08/31/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/29/22 09:42



12423 NE Whitaker Way Portland OR 97230 p.503-254-1794

**Cannabis Chain of Custody Record**

ORELAP ID: OR100028

Analysis Requested												Purchase Order Number:					
												Project Number:					
												Project Name:					
												<input type="checkbox"/> Report Instructions: <input type="checkbox"/> Send to State - METRC <input checked="" type="checkbox"/> Email Final Results: <input type="checkbox"/> Fax Final Results <input type="checkbox"/> Cash/Check/CC/Net 30					
												Other:					
Field ID	Date/Time Collected	Pesticides - OR 59 compounds	Pesticide Multi-Residue - 379 compounds	Potency	Residual Solvents	Water Activity	Moisture	Terpenes	Micro: Yeast and Mold	Micro: E.Coli and Total Coliform	Heavy Metals	Mycotoxins	Other:	Matrix	Weight	Serving size for edibles	Comments/Metric ID
PROC-EG30-CBD	8/24	1								1				isolate			

Collected By:	[Redacted]												<b>Lab Use Only:</b>	
<input type="checkbox"/> Standard (5 day) <input type="checkbox"/> Rush (3-4 day) (1.5x Standard) <input checked="" type="checkbox"/> Priority Rush (2 day) (2x Standard)													Client Alias:	
													Order Number:	
													Proper Container	
													Sample Condition	
													Temperature:	
													Shipped Via:	
													Evidence of cooling: <input type="checkbox"/> Yes <input type="checkbox"/> No	

SUBMISSION OF SAMPLES WITH TESTING REQUIREMENTS TO PIXIS WILL BE UNDERSTOOD TO BE AN AGREEMENT FOR SERVICES IN ACCORDANCE WITH THE CONDITIONS LISTED ON THE BACK OF THIS FORM  
 Revision: 1.02 Control#: CF023 [www.pixislabs.com](http://www.pixislabs.com)  
 Effective 01/31/2019 Revised 01/31/2019 Page 1 of 2



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 22-010282/D002.R000  
**Report Date:** 08/31/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/29/22 09:42



Document ID: 3177 Revision: 3  
Effective: 04/26/2022  
Page 1 of 1

**PACKAGE RECEIVING FORM**

**Delivery Date:** 8/29  Same as Opened By Date  Unsure

How was the package delivered?

UPS FEDEX  USPS DHL OTHER: \_\_\_\_\_

Tracking Number: 9400 1112 0255 7996 6858 37

		CIRCLE ONE	
1) Was package sealed with no evidence of holes/tampering?		<input checked="" type="radio"/> YES	<input type="radio"/> NO
Further custody seal/tampering notes: _____			
2) Was packing material used?		<input type="radio"/> YES	<input checked="" type="radio"/> NO
If YES: <input type="checkbox"/> PEANUTS <input type="checkbox"/> BUBBLE <input type="checkbox"/> WRAP <input type="checkbox"/> FOAM PAPER			
3) Was a Complete Chain of Custody (COC) received?		<input checked="" type="radio"/> YES	<input type="radio"/> NO
Comment (PT?, Email?): _____			
4) Sample temperature upon arrival?			<u>21.2</u> °C
5) Evidence of cooling?		<input type="radio"/> YES	<input checked="" type="radio"/> NO
If YES, What kind? <input type="checkbox"/> ICE <input type="checkbox"/> FREEZER PACK <input type="checkbox"/> DRY ICE			
Insulation? <input type="checkbox"/> PLASTIC COOLER <input type="checkbox"/> STYROFOAM <input type="checkbox"/> OTHER: _____			
6) Were sample containers sealed in separate plastic bags/secondary containment?		<input checked="" type="radio"/> YES	<input checked="" type="radio"/> NO <i>JF</i>
7) Did sample containers arrive in good condition?		<input checked="" type="radio"/> YES	<input type="radio"/> NO
If NO: <input type="checkbox"/> LEAKED <input type="checkbox"/> BROKEN <input type="checkbox"/> OTHER: _____			
If NO: Suspect contamination of other samples? <input type="checkbox"/> YES <input type="checkbox"/> NO			
8) Sample labels present?		<input checked="" type="radio"/> YES	<input type="radio"/> NO
9) Do sample labels agree with COC?		<input checked="" type="radio"/> YES	<input type="radio"/> NO
If NO, number of sample containers received: _____			

**Sample pre-log location:**

R39 R44 F44 R99  CANNA SHELF  FOOD SHELF Other: \_\_\_\_\_

**Other Notes:**

Received By (initials): JF Date: 8/29 Time: 9:44



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 22-010282/D002.R000  
**Report Date:** 08/31/2022  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 08/29/22 09:42





Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitation level raised due to matrix interference.
B	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.